



EUROPEAN UNION

DELEGATION TO THE REPUBLIC OF SERBIA

Belgrade, 03/11/2010

CONTRACTING AUTHORITY'S CLARIFICATIONS No. 2

**Supply of equipment for the Implementation of the National screening programme
for colorectal, cervical and breast cancer**

Publication ref.: EuropeAid/129567/C/SUP/RS

Tender no: 09SER01/05/21

No	Question	Answer
1.	<p><u>Item 7.1 Laboratory Light Microscope with Digital Camera:</u></p> <p>Regarding the item 7.1 (Laboratory Light Microscope with Digital Camera), we would like to clarify that the microscope systems with LED illumination (with a life time of 100,000 hours) are equipped with 5 fold objective turret. Since LED system requires a special type of microscope body, working with more than 5 objectives is not recommended. So, we kindly request you revise “Objective turret for 5 objectives” instead of; “Objective turret for 6 objectives”</p>	<p>This is not acceptable.</p> <p>Please refer to technical specifications of the Tender dossier</p>
2.	Corrigendum	<p>A corrigendum extending the submission deadline to 18/11/2010 will be published on the EuropeAid Website in the coming days</p>