

**CORRIGENDUM NO. 1 to the TENDER DOSSIER**

**Publication Ref: : EuropeAid/129567/C/SUP/RS  
Tender no: (09SER01/05/22)**

**“Implementation of the National screening programme for colorectal, cervical and breast cancer – Supply tender- Re-launch”  
Location - Serbia**

**Instructions to Tenderers**

**The former text**

**2 Timetable**

	DATE	TIME*
<b>Deadline for request for any clarifications from the Contracting Authority</b>	<b>06 June 2011</b>	<b>12:00 CET</b>
<b>Last date on which clarifications are issued by the Contracting Authority</b>	<b>16 June 2011</b>	-
<b>Deadline for submission of tenders</b>	<b>27 June 2011</b>	<b>12:00 CET</b>
<b>Tender opening session</b>	<b>27 June 2011</b>	<b>15:00 CET</b>
<b>Notification of award to the successful tenderer</b>	<b>July 2011</b>	-
<b>Signature of the contract</b>	<b>August 2011</b>	-

\* All times are in the time zone of the country of the Contracting Authority  
• Provisional date

**Shall read as new text:**

**2 Timetable**

	DATE	TIME*
<b>Deadline for request for any clarifications from the Contracting Authority</b>	<b>06 June 2011</b>	<b>12:00 CET</b>
<b>Last date on which clarifications are issued by the Contracting Authority</b>	<b>16 June 2011</b>	-
<b>Deadline for submission of tenders</b>	<b>11 July 2011</b>	<b>12:00 CET</b>
<b>Tender opening session</b>	<b>11 July 2011</b>	<b>15:00 CET</b>
<b>Notification of award to the successful tenderer</b>	<b>August 2011</b>	-
<b>Signature of the contract</b>	<b>August 2011</b>	-

\* All times are in the time zone of the country of the Contracting Authority  
• Provisional date

## ANNEX II Technical Specifications

<b>LOT 1 DIGITAL MAMMOGRAPHY UNITS</b>
<b>Item 1.1 Digital mammography units</b>
<b>The former text</b>
<ul style="list-style-type: none"><li>• Ambient temperature: 40°C (hot zone)</li><li>• Focal spots: 2 (two) small 0,1mm, large 0.3mm</li><li>• 1289x1024 LCD color display, 1 unit</li></ul>
<b>Shall read as new text:</b>
<ul style="list-style-type: none"><li>• Ambient temperature: 35°C -40°C (hot zone)</li><li>• Focal spots: at least 2 (two) small 0,1mm, large 0.3mm</li><li>• Compression plates: 24x30cm and 18x24cm</li><li>• Control (operated) console with integrated lead glass protection to be used in the same room where the system is installed</li><li>• 1280x1024 LCD color display, 1 unit</li></ul>
<b>LOT 4: CYTOPATHOLOGY LABORATORY EQUIPMENT</b>
<b>Item 4.1 Laboratory Light Microscope with Digital Camera</b>
<b>The former text</b>
<ul style="list-style-type: none"><li>• Light microscope with long-life LED illumination (100000h)</li><li>• Ceramic coated movable table, right/left hand use, changeable by user</li><li>• Objective turret for 5-6 objectives</li><li>• Standard condenser, color coding, 2,5x – 100x</li><li>• Brightness synchronized objectives: 2, 2,5x/0.07, 5x/0.12, 10x/0.25, 20x/0.40, 40x/0.65, 100x1.25 oil</li><li>• Digital camera, min.resolution 7 Mpixel (3072x2304)</li><li>• Live image in full size 2048x1536, min. 12 frames/sec.</li><li>• Pixel size 3.2 x 3.2</li></ul>
<b>Shall read as new text:</b>
<ul style="list-style-type: none"><li>• Light microscope with long-life LED illumination (10,000 hours)</li><li>• Ceramic coated movable table. A hard coat anodized surface with right drive shall also be accepted.</li><li>• Turret for a minimum of 5 objectives or more</li><li>• Standard condenser, any type of coding 2,5x – 100x</li><li>• Brightness synchronized objectives: 2, 2x or 2,5x/0.07, 4x or 5x/0.12, 10x/0.25, 20x/0.40, 40x/0.65, 100x1.25 oil</li><li>• Digital camera, min.resolution 5 Mpixel</li><li>• Live image in full size min 1920x1080, min. 12 frames/sec</li><li>• Pixel size min. 3.2 x 3.2</li></ul>
<b>LOT 4: CYTOPATHOLOGY LABORATORY EQUIPMENT</b>
<b>Item 4.3 Laboratory Light Microscope</b>
<b>The former text</b>

- Light microscope with long-life LED illumination (100000h)
- Ceramic coated movable table, right/left hand use, changeable by user
- Turret for 5 objectives
- Standard condenser colour coding 2,5x – 100x
- Brightness synchronized objectives: 2,5x/0.07, 10x/0.25, 20x/0.40, 40x/0.65, 100x/1.25 oil

**Shall read as new text:**

- Light microscope with long-life LED illumination (10,000 hours)
- Ceramic coated movable table. A hard coat anodized surface with right drive shall also be accepted.
- Turret for a minimum of 5 objectives or more
- Standard condenser, any type of coding 2,5x – 100x
- Brightness synchronized objectives: 2, 2x or 2,5x/0.07, 4x or 5x/0.12, 10x/0.25, 20x/0.40, 40x/0.65, 100x1.25 oil

**LOT 4: CYTOPATHOLOGY LABORATORY EQUIPMENT**

**Item 4.5 Autostainer**

**The former text**

- Linear
- Slide basket and hanger: stainless steel

**Shall read as new text:**

- Linear or carrousel model
- Slide hanger: stainless steel, slide basket: stainless steel or plastic.

**All other terms and conditions of the Instructions to tenderers remain unchanged. The above alterations and/or corrections to the Instructions to tenderers are integral part of the Tender Dossier.**

**The amended ANNEX II + III: Technical Specifications + Technical Offer and the amended Appendixes to Technical Specifications are attached to this CORRIGENDUM n.1.**